

A Complete House of Scientific & Laboratory Equipments

LCR / AC BRIDGE - GUPTA BRAND 200664



Technical Specifications:

Inductance Measurements :

Min :	0.2 microH
Max :	110 H or more
Ranges :	8
Accuracy :	+ 1% for greater 'Q' coils

Capacitance Measurements :

Min :	1 pF
Max :	1110 micro Farad or more
Ranges :	8
Accuracy :	+ 1%

Resistance Measurements:

Min :	100 milli Ohms
Max :	10 Megohms
Ranges :	8
Accuracy :	+ 1%
Display :	Numeric 3.5 digits
D/QFactor :	Available for Fine adjustment Q: 0-30 D : 0 to 3
Excitation :	Internal 1KHz + 10% sine wave and DC Power Supply : 220V +/- 10%, 50Hz.

Main Address :

Gupta Agencies

5341, Punjabi Mohalla
Ambala Cantt - 133001
Haryana - INDIA
Mr Ashok Gupta +91-9896301766
Mr Sushil Gupta +91-9896311766
Mr Dixit Gupta +91-9416021766
Landlines - +91-171-4006829, 4006826,
4005829

E-mail:

info@guptaagenciesindia.com
dixit@guptaagenciesindia.com



Website :

www.guptaagenciesindia.com

Manufacturing Unit & Export House :

GUPTA'S ARCADE

Kahsra No 22/1/2/3

Pallavi Enclave

Near Industrial Area

Ambala Cantt - 133001

Haryana -INDIA

Landlines - +91-171-4006829, 4006826,
4005829