

**A Complete House of Scientific & Laboratory Equipments**

**LCR / AC BRIDGE - GUPTA BRAND 200664**



**Technical Specifications:**

**Inductance Measurements :**

Min :	0.2 microH
Max :	110 H or more
Ranges :	8
Accuracy :	+ 1% for greater 'Q' coils

**Capacitance Measurements :**

Min :	1 pF
Max :	1110 micro Farad or more
Ranges :	8
Accuracy :	+ 1%

**Resistance Measurements:**

Min :	100 milli Ohms
Max :	10 Megohms
Ranges :	8
Accuracy :	+ 1%
Display :	Numeric 3.5 digits
D/QFactor :	Available for Fine adjustment Q: 0-30 D : 0 to 3
Excitation :	Internal 1KHz + 10% sine wave and DC Power Supply : 220V +/- 10%, 50Hz.

**Main Address :**

**Gupta Agencies**

5341, Punjabi Mohalla  
Ambala Cantt - 133001  
Haryana - INDIA  
Mr Ashok Gupta +91-9896301766  
Mr Sushil Gupta +91-9896311766  
Mr Dixit Gupta +91-9416021766  
Landlines - +91-171-4006829, 4006826,  
4005829

**E-mail:**

info@guptaagenciesindia.com  
dixit@guptaagenciesindia.com



**Website :**

[www.guptaagenciesindia.com](http://www.guptaagenciesindia.com)

**Manufacturing Unit & Export House :**

**GUPTA'S ARCADE**

Kahsra No 22/1/2/3

Pallavi Enclave

Near Industrial Area

Ambala Cantt - 133001

Haryana -INDIA

Landlines - +91-171-4006829, 4006826,  
4005829